

**TENDER NOTICE FOR CONVERSION OF DELIVERY VAN
INTO AMBULANCE/ RETROFITTING**

Consulate General of India, Jeddah, invites sealed quotations from reputed companies which undertake conversion of delivery vans into ambulances as per the specimen given by the Saudi Red Crescent Society (copy enclosed). The quotations should be submitted to Assistant Vice Consul (Haj), Consulate General of India, Jeddah latest by Wednesday, **31st March 2021 (till 1700 hrs)**.

2. Technical and Financial bids must be submitted in two separate sealed envelopes with "Technical Bid" & "Financial Bid" clearly mentioned on the envelopes. Before submission of technical and financial bids, the interested bidders are advised to attend a pre-bid meeting at Consulate General of India, Jeddah on Sunday, **21st March, 2021 (14:30 hours)**. They are also advised to go through this tender notice in detail to apprise themselves of the information relevant to submission of their bids.

3. **Technical bid:** The interested companies should submit a technical bid (in a separate sealed envelope) consisting of following documents /details:

- Copy of valid license and Registration
- Profile of the company, also reflecting experience in the field conversion and retrofitting

4. **Financial bid:**

i) The interested companies should submit a financial bid (in a separate sealed envelope) giving the **quoted rate per vehicle complete retrofitting of the vehicle to fulfill all requirements of Saudi Red Crescent Society { More details at para 6 (ii) }**.

5. **Opening of bids:**

a) A Tender Evaluation Committee of the Consulate will open the bids on Sunday, 4th April 2021 (0930 hrs) at the Consulate. Representatives of the bidders may like to attend the bid opening.

6. CONDITIONS:

i) The successful bidders should be able to convert/retrofit the ambulance as per the requirement of the Saudi Red Crescent Society and the Consulate.

ii) An indicative list of requirements and fittings to be done by the successful bidder is given at Annexure-I. Additionally, any other requirement of Saudi Red Crescent Society, on the date of delivery of the retrofitted vehicle to the Consulate by the company, will be carried out by the successful bidder without additional cost. Failure to do in a reasonable time provided by the Consulate will lead to cancellation of the order and stopping any further payment as well as blacklisting the company.

iii) Refusing to convert/retrofit particular instruments by a successful bidder at the rate by it will lead to blacklisting of the company and non-payment of any dues to be paid by the Consulate (if any).

7. **PAYMENT**: Payment to the companies (successful bidders) for conversion/retrofit will be made either through bank transfer or cheque after the supply of the items demanded.

8. Consulate General of India, Jeddah, reserves the right to reject this tender without assigning any reason thereof.

CONSULATE GENERAL OF INDIA, JEDDAH

TENDER NOTICE

Quotations are invited from reputed companies which undertakes conversion of vans into ambulances as per the specimen given by the Saudi Red Crescent Society (copy enclosed).

2. Interested companies which undertakes conversion of vans into ambulance can submit their written proposal in a sealed envelope to Assistant Vice Consul(Hajj), Consulate General of India, Tahalia Street, Jeddah latest by **31st March, 2021 (till 17:00 hrs)**. More details are available at the website of the Consulate General of India, Jeddah at

www.cgijeddah.gov.in- 'Haj News'
'Quotation for vehicles with conversion'

Kingdom of Saudi Arabia,
Ministry of Foreign Affairs,
Makkah Al Mukarramah Region Branch

No. 21/3/33952

Date: 21.6.1434H (1.5.2013G)

Most immediate

The Ministry of Foreign Affairs (Makkah Al Mukarramah Region Branch) presents its compliments to the Diplomatic and Consular Missions as well as Offices of the International Organizations accredited in Makkah Al Mukarramah Region.

This is in continuation of the Ministry's circular No. 21/3/21365 dated 27.11.1433H containing regularization of ambulance vehicles of the offices of Haj Affairs of each country and need to abide by these regulations.

The Ministry is pleased to forward herewith the circular of the Ministry of Haj to all authorities concerned to Haj affairs in the countries from where Hajis come containing conditions for the permit of ambulances of temporary Medical Offices of Foreign Hajis and which are supposed to be used within the Haj cities in addition to the equipments required to be available with these vehicles and necessary equipments for the medic bag and oxygen bag as well.

The Ministry requests the Consulate to take note of it and abide by the instructions, requirements and preconditions approved by the Saudi Red Crescent Authority.

The Ministry avails itself of this opportunity to express its best compliments to them.

Kingdom of Saudi Arabia,
Saudi Red Crescent Authority,
General Adm. Makkah
Haj and Umrah Department

Conditions for the permit of ambulances of Medical Offices of Foreign Hajis

1. Letter from Ministry of Haj
2. The driver should be Saudi national fulfilling legal conditions of driving.
3. Vehicle manufacturing date is new which should not to be more than three years old and should not have run more than 50,000 kilometers.
4. Minimum limit for removing the cabin (150 cm width * 320 cm length) to render best ambulance services.
5. Colour of vehicle would be (white with purple colour lines on that with the width of 20 cm) and emblem of the Medical offices would be on both sides as well as on front and back sides.
6. Equipped with alarm and cautioning lights.
7. Passed in the periodic inspection.
8. Equipped by medical instruments (general equipment) as per the attached list
9. To be used for the transportation of injured and patients in the Haj cities (Makkah Al Mukarramah and Madinah Munawwarah) and the ways leading to these places. It is prohibited to use the ambulance vehicle in the Holy Mashair region (Aftaft, Muzdalifa and Mina).
10. A staff capable to undertake relief work is working on the vehicle provided that the level of service providers should not be less than nurse or ambulance technician.
11. Furnishing the concerned authorities with the telephone numbers of the Mission and details of the employees working on that vehicle among the medical service providers and locations of the vehicles with the permit request.
12. Providing the quick maintenance service.

Attachment with details of equipments, medical items and materials for
preparing ambulance can (as per the specifications of Saudi Red
Crescent)


S.No.	Sustainable inventory	S.No.	Consumed inventory
1	Oxygen cylinder -Big	28	Lawn square 4X4
2	Oxygen cylinder -Small	29	Lawn roll 4X8
3	Moving bed	30	Scanners
4	Folding stretcher	31	Oxygen masks all sizes
5	Scoop stretcher	32	Burns cloth
6	Stair chair	33	Ordinary cloth
7	Splint	34	Blanket
8	Tight splint	35	Aluminum blanket
9	Oxygen cylinder-big- regulator	36	Intubation kit
10	Oxygen cylinder -small- regulator	37	Oxygen tubes
11	Blood-Oxygen ratio measuring device	38	Airway (mouth breathing)
12	Sugar- Oxygen ratio measuring device	39	Airway (Nose breathing)
13	Pressure measuring device for adults	40	Renal basins
14	Pressure measuring device for minors	41	Wooden splints
15	Shocks device (self)	42	Neck splints
16	Scout check	43	Intravenous fluids
17	Rings cutter	44	Distillation tool
18	Glass driller	45	Kanveula all sizes
19	Fridge	46	Saranjat all sizes
20	Limbs splint	47	Abrokhz
21	Shock device	48	Slices for sugar measuring
22	Static suction device	49	Lawn twofold
23	Mobile suction device	50	Suction tubes
24	Plastic sheet	51	Oxygen moisturizer
25	Medic bag	52	Gloves
26	Oxygen bag	53	Paper masks
27	Search light	54	Delivery kit
28		55	Paper roll

Contents of medic bag

1	Plus hexameter	13	Injury scissor
2	Sugar device	14	Lawn
3	Oral duct	15	Pressing band
4	Rhinal duct	16	Thermometer
5	Tongue depressor	17	Plaster
6	Ring cutter	18	Burns cloth
7	Glass driller	19	Liquid D5 ranger-normal
8	Twofold bandage	20	Venus catheters
9	Lawn square	21	Swabs
10	Gloves	22	Pressure device big and small
11	Paper mask	23	Plastic goggles
12	Search light	24	Stethoscope

Contents of Oxygen bag

1	Oxygen tube
2	Different Oxygen masks
3	Intubation tube
4	Intubation tubes with different sizes
5	Duct (Oral +rhinal)
6	Intubation fixing tool



عاجل جداً

٥/٥/٢٠١٣

تهدي وزارة الخارجية (فرع منطقة مكة المكرمة) أطيب تحياتها
إلى البعثات الدبلوماسية والقنصليات ومكاتب المنظمات الدولية
المعتمدة في منطقة مكة المكرمة.

إحاقاً لتعميم الوزارة رقم ٢١/٣/٢١٣٦٥ وتاريخ
١٤٣٣/١١/٢٧ هـ المتضمن تنظيم سيارات الإسعاف التابعة لمكاتب
شؤون الحج لكل دولة وضرورة الالتزام بتلك التنظيمات.

يسرها أن تبعث لها بطيه تعميم وزارة الحج إلى عموم الجهات
المعنية بشؤون الحج في الدول التي يقدم منها الحجاج المتضمن شروط
طلب التصريح لسيارات الإسعاف التابعة للمكاتب الطبية المؤقتة للحجاج
الأجانب والمراد استخدامها داخل مدن الحج بالإضافة للتجهيزات الواجب
توفرها بتلك المركبات و التجهيزات الضرورية لحقيبة المسعف و حقيبة
الأوكسجين.

تأمل الوزارة من القنصلية الاطلاع و الالتزام بالتعليمات والضوابط
والاشتراطات المعتمدة من قبل هيئة الهلال الأحمر السعودي.



وتنتهز الوزارة هذه الفرصة لتعرب لها عن أطيب تحياتها



شروط التصريح بسيارات إسعاف المكاتب الطبية للحجاج الأجانب

1. خطاب من وزارة الحج.
2. أن يكون قائد العربة سعودي الجنسية مستوفي الشروط النظامية للقيادة.
3. تاريخ صنع السيارة حديث لا يزيد عن 3 سنوات ولا تكون قطعت أكثر من 50.000 ألف كم.
4. الحد الأدنى لإبعاد الكينة (150 سم عرض * 220 سم طول) لتقديم أفضل الخدمات الإسعافية.
5. لون السيارة (ابيض وعليها خط بنفسجي (مرفق غامق) بعرض 20 سم) وشعار المكاتب الطبية للحجاج على الجانبين ومن الأمام والخلف للإسعاف.
6. تجهزه بمنبه صوتي وإضاءات تحذيرية.
7. تجتاز الفحص الدوري.
8. تجهزه طبيًا (تجهيز عام) حسب البيان المرفق.
9. تستخدم فقط لنقل الإصابات والحالات المرضية في مدن الحج (مكة المكرمة والمدينة المنورة) والطرق المؤدية إليها ويتم استخدامها داخل المشاعر المقدسة (عرفات ومزدلفة ومنى).
10. يعمل عليها طاقم مؤهل للعمل الاسعافي على أن لا يقل مستوى مقدمي الخدمة عن ممرض أو فني إسعاف.
11. تزويد الجهات المختصة بموافقات البعثة وبيانات العاملين على السيارة من مقدمي الخدمة الطبية ومواقع تواجد السيارات مع طلب الترخيص.
12. توفير خدمة الصيانة السريعة.

الشفوعات : / / التاريخ :
Enclosures : Date : / / No :

تلفون: ٠٢ ٦٩٣٧٤٦٣ - فاكس: ٠٢ ٦٩٣٧١٩٧ - مكة المكرمة

Tel.: 02 6937463 - Fax : 02 6937197 - Makkah Al-Mukarramah - www.srca.org.sa - E-mail: mak@srca.org.sa

Kingdom of Saudi Arabia
Saudi Red Crescent Authority
General Adm. Makkah
Hajj and Omrah Department



الهيئة العامة للصليب الأحمر السعودي
هيئة الهلال الأحمر السعودي
الإدارة العامة بمنطقة مكة المكرمة
أمانة التفتيش والتفتيش

بيان متطلبات الجودة الإستراتيجية

رقم	الخدمة المستهدفة	الخدمة المستهدفة
١	استطوانة أكسجين كبير	شاش مربع ٤*٤
٢	استطوانة أكسجين صغيرة	شاش لفر ٨*٤
٣	المنبر المتحرك	ماسحات كحول
٤	لقالة قابلة للطي	ماسكات أكسجين جميع المقاسات
٥	لقالة سكوب	شرشف حمولي
٦	كرسي الدرج	شرشف عادي
٧	بجيرة kid	بطانية
٨	بجيرة ضد	بطانية المنيوم
٩	منظم أكسجين كبير	طقم تسيب
١٠	منظم أكسجين صغير	النايب أكسجين
١١	جهاز قياس نسبة الأكسجين بالدم	أبروي (بجري الهواء القسوي)
١٢	جهاز قياس نسبة السكر بالدم	أبروي (بجري الهواء الأنفي)
١٣	جهاز قياس الضغط للكبار	أحواض كلوية
١٤	جهاز قياس الضغط للأطفال	جباير خشبية
١٥	جهاز صدعات (ذاني)	جباير عنقية
١٦	كشف فحص	شماليل زويدية
١٧	قاطع خواتم	أداة تقطير
١٨	تألب زجاج	كانبولة جميع المقاسات
١٩	ذلاجة	سرنجات جميع المقاسات
٢٠	جباير أطراف (لريك بالك)	حص ابو رنجر
٢١	جهاز صدعات	شرائح لقياس السكر
٢٢	جهاز شفط ثابت	شاش مضاعف
٢٣	جهاز شفط متحرك	النايب شفط
٢٤	لوح بلاستيكي	مرطب أكسجين
٢٥	شظية مسعف	قلنرات
٢٦	شظية أوكسجين	ماسكات ورقية
٢٧	كشف	طقم ولادة
٢٨		رول ورقي

الرقم : التاريخ : / /
Endosures : Date : / / No :

تلفون: ٢٦٩٣٧٤٦٣ فاكس: ٢٦٩٣٧١٩٧ مكة المكرمة
Tel.: 02 6937463 - Fax: 02 6937197 - Makkah Al Makkah



مستلزمات مختبر الأوكسجين

مقص إصابات	١٣	باجس أكسفوجين	١
شاش	١٤	جهاز بكر	٢
رباط ضاغط	١٥	مجري هواء فمى	٣
مقياس حرارة	١٦	مجري هواء الفمى	٤
بلمستر	١٧	خافض لسان	٥
ششفخ تجرؤق	١٨	قاطع خواتم	٦
محلول D5 - رنجير - نورمل	١٩	ناقب زجاج	٧
قساطير وريدية	٢٠	ضماد مضاعف	٨
مسحات	٢١	شاش مربع	٩
جهاز ضغط كبير وصغير	٢٢	قفزات	١٠
نظارة واقية	٢٣	كمام وجه ورقى	١١
بماعة طبيب	٢٤	كشاف لحيص	١٢

مستلزمات حقيبة الأوكسجين

أنبوبة أكسجين مع منضمها	١
أقنة أكسجين مختلفة	٢
جهاز تنبيب	٣
أنابيب تنبيب بمقاسات مختلفة	٤
مجري هواء (فمى + أنفى)	٥
أداة تثبيت التنبيب	٦

الرقم : التاريخ : / / المشفوعات :
Endosures : Date : / / No :

تلفون: ٠٢ ٦٩٣٧٤٦٣ - فاكس: ٠٢ ٦٩٣٧١٩٧ - مكة المكرمة

Tel.: 02 6937463 - Fax: 02 6937197 - Makkah Al-Mukramah - www.srca.org.sa - E-mail: mak@srca.org.sa